

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control Low 5x1mL

REF 5006330
LOT 2B61B02.01
 30-Jun-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2B61B02	30-Jun-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F61B00	75 - 212	189.3 OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D62B00	42 - 103	66.8 OK
	TGA Reagenz RD (nm Peak Thrombin) LOT 7G61B00	228 - 347	249.0 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

08-jul-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

08-jun-2016
Datum / Date: