

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0300.01
---	---	-----------

Technoview® Rivaroxaban CAL Set High 5x1mL

REF 5090171
LOT 0361B00.01
 31-Jan-2017

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Technoview Rivaroxaban CAL1 High	5161B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL2 High	5261B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL3 High	5361B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL4 High	5411B02	31-Jan-2017
1x1mL	Technoview Rivaroxaban CAL5 High	5561B00	31-Jan-2018


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control High LOT 3351B00	242.6 – 404.3 ng/mL	330.5	OK
	Rivaroxaban Control Medium LOT 3551B00	114.9 – 191.5 ng/mL	174.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

13. MAI 2016
Datum / Date:

Freigabe von
Approved by


Fikret NASUFI

13. MAI 2016
Datum / Date: