
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3400.01 <b>LOT</b> 3453B00.01
---	---	------------------------------------

**TECHNOVIEW® Rivaroxaban Control Low 5x1mL**

**REF**            5090172  
**LOT**            3453B00.01  
        30-Nov-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		41.6 – 77.2 ng/mL	59.4	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	41.6 – 77.2 ng/mL	57.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
**Lucas ROSENMAYR**

10. FEB. 2016  
Datum / Date:

**Freigabe von**  
**Approved by**

  
**Christoph ENNE**

10. FEB. 2016  
Datum / Date: