



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1000.01

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269

LOT 1062B00.02

⌚ 28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	TECHNOVIEW Apixaban Cal 1	1161B00	28-Feb-2018
1x1mL	TECHNOVIEW Apixaban Cal 2	1261B00	31-Mar-2018
1x1mL	TECHNOVIEW Apixaban Cal 3	1361B00	30-Apr-2018
1x1mL	TECHNOVIEW Apixaban Cal 4	1461B00	28-Feb-2018
1x1mL	TECHNOVIEW Apixaban Cal 5	1561B00	31-Mar-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont low LOT3943B01	114.1 – 211.9 ng/mL	146.1 OK
	TECHNOVIEW Apixaban Cont high LOT3D43B01	282.3 – 470.5 ng/mL	316.8 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von  
Prepared by

Christoph ENNE

01. JULI 2016

Datum / Date:

Freigabe von  
Approved by

Lavdrim MASUFI

01. JULI 2016

Datum / Date: