



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRO00AZ

TECHNOZYM® ADAMTS-13 ELISA Kit 2x48T.


REF 5450501

LOT RQ63B00.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
2x48T	Adamts13 Plate	RQS61B0	30-Apr-2018
1x0.3mL	Adamts13 Conjugate	RQV32B2	30-Apr-2018
2x3mL	AdamTS13 Activity Substrate	RQW61B0	30-Apr-2018
1x6mL	AdamTS13 Antigen Substrate	RQX33B2	30-Apr-2018
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL33B2	30-Apr-2018
1x6mL	AdamTS13 Antigen Stop Solution	RQK33B2	30-Apr-2018
1x80mL	Wash buffer Concentrate	RNA62B0	30-Jun-2018
1x90mL	Incubation buffer	RNB62B0	31-May-2018
1x0.5mL	Adamts 13 Calibrator 1	1UT62B0	30-Jun-2018
1x0.5mL	Adamts 13 Calibrator 2	2UT61B0	31-May-2018
1x0.5mL	Adamts 13 Calibrator 3	3UT61B0	31-May-2018
1x0.5mL	Adamts 13 Calibrator 4	4UT62B0	31-May-2018
1x0.5mL	Adamts 13 Calibrator 6	6UT62B0	31-May-2018
1x0.5mL	Adamts 13 High Control	RUY61B0	31-May-2018
1x0.5mL	Adamts 13 Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY61B0 activity	0.50 – 0.84 IU/mL	0.65	OK
	High control RUY61B0 antigen	0.46 – 0.76 IU/mL	0.54	OK
	Low control RUZ61B0 activity	0.10 – 0.24 IU/mL	0.23	OK
	Low control RUZ61B0 antigen	0.11 – 0.27 IU/mL	0.27	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥200	441.0	OK
	Gain: 59 Antigen ΔRFU Cal1-Cal5	≥200	230	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK


	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQRO00AZ</p>
---	---	---

Durchführung von
Prepared by


Michael KAFKA

15-Jul-2016
 Datum / Date:

Freigabe von
Approved by


Christoph ENNE

15-Jul-2016
 Datum / Date: