

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQQT00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------	----------

TECHNOZYM ADAMTS-13 Control Set 2x0.5mL

REF 5450563
LOT 7QT61B0.01
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
1x0.5ml	Adamts 13 High Control	RUY61B0	31-May-2018
1x0.5ml	Adamts 13 Low Control	RUZ61B0	31-May-2018


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability Plate: RQS61B0 Conjugate: RQV32B2	High control RUY61B0 activity	0.50 – 0.84 IU/mL	0.63	OK
	High control RUY61B0 antigen	0.46 – 0.76 IU/mL	0.55	OK
	Low control RUZ61B0 activity	0.10 – 0.24 IU/mL	0.18	OK
	Low control RUZ61B0 antigen	0.11 – 0.27 IU/mL	0.18	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

**Durchführung von
Prepared by**

Michael KAFKA 

15-Me-2016
Datum / Date:

**Freigabe von
Approved by**

Christoph ENNE 

15-Jul-2016
Datum / Date: