



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601  
LOT RR63B00.01  
⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 Ag Plate	RRS61B0	30-Apr-2018
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B2	30-Apr-2018
1x 80mL	Wash buffer Concentrate	RNA62B0	30-Jun-2018
1 x 90mL	Incubation buffer	RNB62B0	31-May-2018
1 x 12mL	TMB Substrate	RKN61B0	31-May-2018
1 x 12mL	Stop solution	RKK61B0	31-May-2018
1 x 0.5mL	Adamts13 Ag Calibrator 1	1UT62B0	30-Jun-2018
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT62B0	31-May-2018
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT62B0	31-May-2018
1 x 0.5mL	Adamts13 Ag High Control	RUY61B0	31-May-2018
1 x 0.5mL	Adamts13 Ag Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.97	OK
Verwendbarkeit Usability	High Control (RUY61B0)	0.43 – 0.71 IU/mL	0.56	OK
	Low Control (RUZ61B0)	0.11 – 0.27 IU/mL	0.18	OK

Durchführung von  
Prepared by

Michael KAFKA 

11-Jul-2016

Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI 

13-Jul-2016

Datum / Date: