

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ8UT0AZ.02
---	---	-------------


**TECHNOZYM® ADAMTS-13 Activity Calibrator Set 6x0.5mL**

REF      5450761  
LOT      8UT61B0.01  
      31-May-2018

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x0.5mL	Adamts-13 Activity Calibrator 1	1UT61B0	31-May-2018
1x0.5mL	Adamts-13 Activity Calibrator 2	2UT61B0	31-May-2018
1x0.5mL	Adamts-13 Activity Calibrator 3	3UT61B0	31-May-2018
1x0.5mL	Adamts-13 Activity Calibrator 4	4UT62B0	31-May-2018
1x0.5mL	Adamts-13 Activity Calibrator 5	5UT61B0	31-May-2018
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.42	OK
	Calibrator 6	< 0.2	0.03	OK
Verwendbarkeit / Usability	High Control (RUY61B0)	0.62 – 1.03 IU/mL	0.89	OK
	Low Control (RUZ61B0)	0.17 – 0.41 IU/mL	0.32	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI 

27-Jun-2016  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA 

27-Jun-2016  
Datum / Date: