




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7UT0AZ.02

TECHNOZYM[®] ADAMTS-13 Activity Control Set 2x0.5mL

REF 5450763

LOT 7UT61B0.01

 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
1x0.5mL	Adamts-13 Activity High Control	RUY61B0	31-May-2018
1x0.5mL	Adamts-13 Activity Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (RUY61B0)	0.62 – 1.03 IU/mL	0.89 OK
	Low Control (RUZ61B0)	0.17 – 0.41 IU/mL	0.32 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

27-Jun-2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

27-Jun-2016

Datum / Date: