




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ0T00AZ

TECHNOTHROMBIN[®] TGA Kit 3x16 T.

REF 5006010

LOT 0T64B00.01

 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P61B00	31-Jul-2018
3x1.5mL	TGA substrat	8F61B00	31-Oct-2017
1x1.5mL	TGA reagent RD	7G62B00	30-Jun-2018
1x0.5mL	TGA reagent RC Low	7D63B00	31-May-2018
1x0.5mL	TGA reagent RC High	7F62B00	28-Feb-2018
1x0.5mL	TGA Thrombin Calibrator	5T62B	31-Mar-2018
1x1mL	TGA Control High	1B63B00	30-Sep-2017
1x1mL	TGA Control Low	2B63B01	30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1B63B00)	140 – 469	379.0	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2B63B01	75 – 212	208.5	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1B63B00)	115 – 359	343.5	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2B63B01	42 - 103	78.4	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1B63B00)	341 – 481	366.6	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2B63B01	228 – 347	234.4	OK



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0T00AZ

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

05. AUG. 2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

08. AUG. 2016
Datum / Date: