



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

TECHNOTHROMBIN® TGA Kit 3x16 T.

REF 5006010

LOT 0T65B00.01

30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P62B00	31-Jul-2018
3x1.5mL	TGA substrat	8F62B00	31-Oct-2017
1x1.5mL	TGA reagent RD	7G63B00	30-Jun-2018
1x0.5mL	TGA reagent RC Low	7D64B00	31-May-2018
1x0.5mL	TGA reagent RC High	7F63B00	28-Feb-2018
1x0.5mL	TGA Thrombin Calibrator	5T63B	31-Mar-2018
1x1mL	TGA Control High	1B64B00	30-Sep-2017
1x1mL	TGA Control Low	2B64B01	30-Sep-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1B64B00)	140 – 469	389.3	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2B64B01	75 – 212	141.5	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1B64B00)	115 – 359	267.7	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2B64B01	42 - 103	53.6	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1B64B00)	341 – 481	390.3	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2B64B01	228 – 347	233.6	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

05. AUG. 2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

08. AUG. 2016  
Datum / Date: