



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRT00AZ

TECHNOZYM[®] ADAMTS-13 Antigen CAL Set 5x0.5mL

REF 5450661
LOT 6RT62B0.01
⌚ 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x0.5mL	HFP: UT ADAMTS-13 Calibrator 1	1UT62B0	30-Jun-2018
1x0.5mL	HFP: UT ADAMTS-13 Calibrator 2	2UT61B0	31-May-2018
1x0.5mL	HFP: UT ADAMTS-13 Calibrator 3	3UT61B0	31-May-2018
1x0.5mL	HFP: UT ADAMTS-13 Calibrator 4	4UT62B0	31-May-2018
1x0.5mL	HFP: UT ADAMTS-13 Calibrator 6	6UT62B0	31-May-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Plate: RRS61B0 Conjugate: RQV32B2	Δ OD Cal 1 – Cal 5	≥ 0.7	0.96	OK
	High Control (RUY61B0)	0.43 – 0.71 IU/mL	0.57	OK
	Low Control (RUZ61B0)	0.11 – 0.27 IU/mL	0.20	OK

Durchführung von
Prepared by

Michael KAFKA

13-JUN-2016
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

13-JUN-2016
Datum / Date: