
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6MBA0
---	---	---------

TECHNOCLOT® PT Owren manual 10x10mL

REF 5005037
LOT 6M64BB0.01
 30-Jun-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Thrombotrack	CCN LOT 1P37B00	70.3 – 201.8 % 0.82 – 1.16 INR	85.4 1.01	OK
	CCAK LOT 2B35B00	13.1 – 19.8 % 2.64 – 3.71 INR	16.1 3.17	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas
Lucas ROSENMAYR

17. AUG. 2016

Datum / Date:

Freigabe von
Approved by

Laydrim
Laydrim NASUFI

19. AUG. 2016

Datum / Date: