



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7G00AZ

TECHNOTHROMBIN[®] TGA RD Reagenz 5x1.5mL

REF 5006220

LOT 7G61B00.01

⌚ 30-Jun-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1.5 mL	TGA RD Reagenz	7G61B00	30-Jun-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00)	341 - 481 443.6 OK
	TGA Control Low (nM Peak Thrombin) LOT 2B61B03)	228 - 347 241.3 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

29. AUG. 2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

29. AUG. 2016
Datum / Date: