

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0D00AZ
---	---	----------

TECHNOTHROMBIN® TGA Cal Set

REF 5006345
LOT 0D63B00.01
 31-Mar-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P31B02	31-Jul-2018
1 x 0.5mL	TGA Thrombin Calibrator	5T61B	31-Mar-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		800-900 nM	841	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.99	OK

Durchführung von
Prepared by

Fikret NASUFI 

29. AUG. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

29. AUG. 2016

Datum / Date: