
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9P00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA BUF 5x3 mL

REF      5006360  
LOT      9P31B02.01  
      31-Jul-2018

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5 x 3 mL	TGA buffer	9P31B02	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / usability  Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.99	OK

Durchführung von  
Prepared by

Fikret NASUFI 

29. AUG. 2016

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA 

29. AUG. 2016

Datum / Date: