



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6V0BAZ.03

Dapttin TC 6x10mL

REF 5035090  
LOT 6V57BB0.01  
⌚ 31-Dez-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		6.6 – 7.2	6.8	OK	
Verwendbarkeit Usability Ceveron® alpha	Dapttin TC, sec	CCN LOT 1P37B00	33.1 – 40.9 sec	36.6	OK
		CCA LOT 3P38B00	96.5 – 119.1 sec	104.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Christoph ENNE

10. MAI 2016

Datum / Date:

Freigabe von  
Approved by

Laydrin NASUFI

10. MAI 2016

Datum / Date: