



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6L0BAZ.01

Siron LIS (aPTT liquid) 10x4 mL

REF 5035119  
LOT 6L62BA0.01  
⌚ 31-Mar-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.0 – 7.6	7.1	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B01	27.5 – 33.9 sec	28.1 OK
	CCA LOT 3P38B01	68.4 – 84.4 sec	81.1 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von  
Prepared by

Stefan STEINER

15. SEP. 2016

Datum / Date:

Freigabe von  
Approved by

Lucas ROSENMAYR

15. SEP. 2016

Datum / Date: