



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2600.01

TECHNOVIEW Arixtra CAL 5x1mL

REF 5090010

LOT 2661B00.01

30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Arixtra Cal 1	2161B00	30-Apr-2018
1x1mL	TECHNOVIEW Arixtra Cal 2	2261B00	30-Apr-2018
1x1mL	TECHNOVIEW Arixtra Cal 3	2361B00	30-Apr-2018
1x1mL	TECHNOVIEW Arixtra Cal 4	2461B00	30-Apr-2018
1x1mL	TECHNOVIEW Arixtra Cal 5	2561B00	30-Apr-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	TECHNOVIEW Arixtra Cont low LOT 3U51B00	0.32 – 0.60µg/mL	0.47	OK
	TECHNOVIEW Arixtra Cont high LOT 3V31B01	0.71 – 1.07µg/mL	0.98	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von  
Prepared by

Christoph ENNE

11. AUG. 2016

Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

11. AUG. 2016

Datum / Date: