



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4900AZ.01

F IX Mangelplasma nativ, 50x1mL

REF 5164016
LOT 4962B00.01
⌚ 31-Jan-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P37B00	80.3 – 133.8 %	119.3	OK
	CCA LOT 3P38B00	15.8 – 26.3 %	23.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

25. FEB. 2016

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR

25. FEB. 2016

Datum / Date: