
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00.01
---	---	-----------

Ceveron Coagulation Reference, 5x1mL

REF 5220110
LOT 1R51B01.01
 28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, Clauss meth. CCN LOT 1P37B00	2.52 – 3.78 g/L	3.24	OK
	Fibrinogen, Clauss meth. CCA LOT 3P38B00	0.93 – 1.39 g/L	1.22	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Laydrin NASUFI

18. APR. 2016

Datum / Date:

**Freigabe von
Approved by**


Lucas ROSENMAYR

18. APR. 2016

Datum / Date: