



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRU00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 Activity ELISA 96T.

REF 5450701  
 LOT RU6DB00.01  
 ⏳ 28-Feb-2018

Inhalt/ Content	Reagenz/Reagent	LOT	⏳
1 x 12	Adamts13 Activity Strips	635	28-Feb-2018
1 x 12mL	Adamts13 Activity HRP Conj antibody	635	28-Feb-2018
1 x 53mL	Wash buffer concentrate	635	28-Feb-2018
1 x 30mL	Reaction buffer	635	28-Feb-2018
1 x 12mL	Color Reagent	635	28-Feb-2018
1 x 12mL	Stop solution	635	28-Feb-2018
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	635	28-Feb-2018
1 x	Sample Dilution MTP	635	28-Feb-2018
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT62B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity High Control	RUY61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.0	OK
	OD Cal 6	< 0.2	0.0	OK
	High Control (RUY61B0)	0.62 – 1.03 IU/mL	0.90	OK
	Low Control (RUZ61B0)	0.17 – 0.41 IU/mL	0.26	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von  
Prepared by

Michael KAFKA

A handwritten signature in black ink, appearing to read "Michael Kafka", written over a horizontal line.

26. SEP. 2016

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI

A handwritten signature in black ink, appearing to read "Fikret Nasufi", written over a horizontal line.

26. SEP. 2016

Datum / Date: