



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7C00AZ

Ceveron[®] TGA RB 5x1 mL

REF 5006202

LOT 7C61B.01

⌚ 30-Sep-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1 mL	Ceveron [®] TGA Reagenz B	7C61B	30-Sep-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B65B00	106 - 177	161.9	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B66B00	77 - 128	104.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

13. OKT. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

13. OKT. 2016

Datum / Date: