
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ.01
---	---	-------------

TECHNOTHROMBIN[®] TGA MFC 5x1ml

REF 5006351
LOT 2B65B00.01
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
5 x 1 mL	TGA MFC 1mL	2B65B00	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	TGA MFC (2B65B00)	PPP: 70-109 MFP: 25-76	PPP: 89 MFP: 45	OK

Durchführung von
Prepared by

Fikret NASUFI 

13. OKT. 2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

13. OKT. 2016
Datum / Date: