



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6L0BAZ.01

Siron LIS (aPTT liquid) 10x10 mL

REF 5035121
LOT 6L61BB0.01
⌚ 31-Mar-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.0 – 7.6	7.1	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	27.5 – 33.9 sec	30.2 OK
	CCA LOT 3P38B01	68.4 – 84.4 sec	82.9 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Christoph ENNE

08. AUG. 2016

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

08. AUG. 2016

Datum / Date: