




**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRB00AZ

**TECHNOZYM® vWF:CBA ELISA Kit 96T**

**REF** 5450301  
**LOT** RB69B00.01  
 30-Apr-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	vWF:CBA Plate Collagen Typ III	RNC64B0	31-Jul-2018
1 x 0.3mL	CBA Conjugate	RNE61B0	31-May-2018
1x 80mL	Wash buffer Concentrate	RNA62B0	30-Jun-2018
1 x 90mL	Incubation buffer	RNB62B0	31-May-2018
1 x 12mL	TMB Substrate	RKN61B0	31-May-2018
1 x 12mL	Stop solution	RKK61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.15	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.63 – 1.05 IU/mL	0.91	OK
	Low Control (BND61B0)	0.13 – 0.29 IU/mL	0.22	OK

Durchführung von  
Prepared by

Fikret NASUFI 

16. AUG. 2016

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA 

16. AUG. 2016

Datum / Date: