



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRE00AZ

TECHNOZYM® PAI-1 Actibind ELISA Kit 96T

REF TC16075
 LOT RE61B00.01
 30-Nov-2017

Inhalt/ Content	Reagenz/Reagent	LOT	⌚
1x	Anti t-PA coated plate Actibind	RGP61B0	30-Jun-2018
1 x 0.3 mL	anti PAI-1 Conjugate	RKQ51B1	31-Jul-2018
1x 80 mL	Wash buffer Concentrate	RNA62B0	30-Jun-2018
1 x 90 mL	Incubation buffer	RNB62B0	31-May-2018
1 x 12 mL	TMB Substrat 12ml	RKN61B0	31-May-2018
1 x 12 mL	Stop solution	RKK61B0	31-May-2018
1 x 0.2 mL	Calibrator 1 active PAI-1	1DW51B0	30-Nov-2017
1 x 0.2 mL	Calibrator 2 active PAI-1	2DW51B0	30-Nov-2017
1 x 0.2 mL	Calibrator 3 active PAI-1	3DW51B0	30-Nov-2017
1 x 0.2 mL	Calibrator 4 active PAI-1	4DW51B0	30-Nov-2017
1 x 0.2 mL	Calibrator 5 active PAI-1	5DW51B0	30-Nov-2017
1 x 0.2 mL	High Control active PAI-1	ADW51B0	30-Nov-2017
1 x 0.2 mL	Low Control active PAI-1	BDW51B0	30-Nov-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	1 – 2.5	1.22	OK
	OD Cal 5	< 0.2	0.05	OK
	High Control (ADW51B0)	27.20 – 45.34 IU/mL	41.51	OK
	Low Control (BDW51B0)	7.55 – 17.63 IU/mL	14.33	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

01. SEP. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

01. SEP. 2016
Datum / Date: