



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQK900AZ.01

TechnoLEIA® D-Dimer Latex Kit 50T.

REF 4847210

LOT K961B00.01

⌚ 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1x4mL	TechnoLEIA® D-Dimer Latex 4mL	K663BA0	30-Sep-2017
1x7mL	TechnoLEIA® D-Dimer Reaktionspuffer 7mL	K563BA0	30-Sep-2017
1x8mL	TechnoLEIA® D-Dimer NaCl 0.9%	K463B00	30-Sep-2021
1x1mL	TechnoLEIA® D-Dimer Calibr. 0 ng/mL	K351B00	30-Nov-2017
1x1mL	TechnoLEIA® D-Dimer Calibr. 3345 ng/mL	K152B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	D-Dimer Control High LOT K241B01	1585.4 – 2642.4ng/mL	2504.3	OK
	D-Dimer Control Low LOT K812B02	210.4 – 350.7ng/mL	340.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas ROSENMAYR

29. SEP. 2016

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

29. SEP. 2016

Datum / Date: