

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQK200AZ.01
---	---	-------------

### TechnoLEIA® D-Dimer Control High 5x1mL

REF	4847230
LOT	K241B01.01
	31-Oct-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.7	OK
Verwendbarkeit D-Dimer Usability D-Dimer Ceveron® alpha	1585.4 – 2642.4 ng/mL	2045.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Vorbereitung von**  
**Prepared by**

*Fikret Nasufi*  
**Fikret NASUFI**

14-Dec-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

**Michael KAFKA** *[Signature]*

14-Dec-2015  
Datum / Date: