

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQK800AZ.01</p>
---	---	--

TechnoLEIA® D-Dimer Control Low 5x1mL

REF 4847232
 LOT K812B02.03
 28-Feb-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit D-Dimer Usability D-Dimer Ceveron® alpha	210.4 – 350.7 ng/mL	266.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Vorbereitung von
Prepared by

Fikret NASUFI 

09-Jun-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

09-Jun-2016
Datum / Date: