

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ5T00AZ
---	---	----------

TECHNOTHROMBIN® TGA Cal Set

REF 5006345
LOT 0D61B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P31B01	30-Apr-2017
1 x 0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
		Kalibrierung / Calibration gegen / against WHO 01/580	800-900 nM
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve ≥ 0.990	0.993	OK

Durchführung von
Prepared by

Michael KAFKA 

25-MAR-2016

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR 

23. MRZ. 2016

Datum / Date: