



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ5T00AZ

TECHNOTHROMBIN[®] TGA Cal Set

REF 5006345

LOT 0D61B00.02

 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P31B01	30-Apr-2017
1 x 0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		800-900 nM	841	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	0.99	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

02-May-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

02-May-2016
Datum / Date: