



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0H00AZ.01

TECHNOVIEW LMW CAL Set 5x1mL

REF 5090040
LOT 0H62B00.01
⌚ 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	TECHNOVIEW LMW Cal 1	5A12B03	30-Jun-2018
1x1mL	TECHNOVIEW LMW Cal 2	5B12B03	30-Jun-2018
1x1mL	TECHNOVIEW LMW Cal 3	5C12B03	30-Jun-2018
1x1mL	TECHNOVIEW LMW Cal 4	5D11B03	30-Jun-2018
1x1mL	TECHNOVIEW LMW Cal 5	5E52B00	31-Dez-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW LMW Cont low LOT 3F51B00	0.25 – 0.46 IU/mL 0.31 OK
	TECHNOVIEW LMW Cont med. LOT 3G51B00	0.59 – 1.09 IU/mL 0.75 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Lucas ROSENMAYR

26. AUG. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA

26. AUG. 2016

Datum / Date: