
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3500AZ.01
---	---	-------------

**TECHNOVIEW Rivaroxaban CONT M 5x1mL**

**REF**            5090173  
**LOT**            3551B00.01  
        30-Nov-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron Alpha®	Anti Xa Rivaroxaban	114.9 – 191.5 ng/mL	145.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von  
Prepared by**

  
**Lucas ROSENMAYR**

**11. DEZ. 2015**  
Datum / Date:

**Freigabe von  
Approved by**

  
**Michael KAFKA**

  
Datum / Date: