



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2K00.02

FVIII Inhibitorplasma HCV neg. 5x1mL

REF 5159010

LOT 2K51B0D.05

⌚ 31-May-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.4	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII Inhibitor clotting, Daptin TC	14.1 – 23.5 BE/mL	14.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Lucas ROSENMAYR

17. AUG. 2015

Datum / Date:

Freigabe von  
Approved by

  
Lavdrim NASUFI

17. AUG. 2016

Datum / Date: