




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4900AZ.01

F IX Mangelplasma nativ, 50x1mL

REF 5164016

LOT 4965B00.01

 31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B01	80.3 – 133.8 %	96.5	OK
	CCA LOT 3P38B01	15.8 – 26.3 %	18.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

16. NOV. 2016
Datum / Date:

Freigabe von
Approved by


Stefan STEINER

16. NOV. 2016
Datum / Date: