

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

TECHNOZYM[®] ADAMTS-13 Activity ELISA 96T.

REF 5450701
LOT RU6EB00.01
 31-Mar-2018

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	640	31-Mar-2018
1 x 12mL	Adamts13 Activity HRP Conj antibody	640	31-Mar-2018
1 x 53mL	Wash buffer concentrate	640	31-Mar-2018
1 x 30mL	Reaction buffer	640	31-Mar-2018
1 x 12mL	Color Reagent	640	31-Mar-2018
1 x 12mL	Stop solution	640	31-Mar-2018
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	640	31-Mar-2018
1 x	Sample Dilution MTP	640	31-Mar-2018
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT62B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity High Control	RUY61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.1	OK
	OD Cal 6	< 0.2	0.0	OK
	High Control (RUY61B0)	0.62 – 1.03 IU/mL	0.85	OK
	Low Control (RUZ61B0)	0.17 – 0.41 IU/mL	0.31	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

Durchführung von
Prepared by

Michael KAFKA 

08. NOV. 2016
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

08. NOV. 2016
Datum / Date: