



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

TECHNOTHROMBIN<sup>®</sup> TGA Kit 3x16 T.

REF 5006010

LOT 0T67B00.01

31-Oct-2017

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P31B02	31-Jul-2018
3x1.5mL	TGA substrat	8F52B00	31-Oct-2017
1x1.5mL	TGA reagent RD	7G61B00	30-Jun-2018
1x0.5mL	TGA reagent RC Low	7D62B00	31-May-2018
1x0.5mL	TGA reagent RC High	7F61B00	28-Feb-2018
1x0.5mL	TGA Thrombin Calibrator	5T61B	31-Mar-2018
1x1mL	TGA Control High	1Z61B00	31-Oct-2017
1x1mL	TGA Control Low	2863B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	407 - 678	553.1	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2863B00	63 - 147	128.3	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	176 - 294	269.9	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2863B00	57 - 143	112.3	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	402 - 670	442.3	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2863B00	365 - 608	477.9	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

24. NOV. 2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

24. NOV. 2016  
Datum / Date: