





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQ8E00AZ

Ceveron[®] TGA SUB 5x3 mL

REF 5006237
LOT 8E61B00.01
 28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	
5x3 mL	Ceveron [®] TGA SUB	8E61B00	28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls LOT 1B52B00	Peak Thrombin nM	TGA RCH f.Ceveron LOT 7F52B	66 - 310	207.4	OK
		TGA RCL f.Ceveron LOT 7D42B	92 - 177	142.1	OK
	Lag time min	TGA RCH f.Ceveron LOT 7F52B	1.7 - 2.8	2.1	OK
		TGA RCL f.Ceveron LOT 7D42B	3.8 - 6.0	4.0	OK
	AUC nM	TGA RCH f.Ceveron LOT 7F52B	1078 - 2845	2049.7	OK
		TGA RCL f.Ceveron LOT 7D42B	1633 - 2500	2027.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

03-May-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

03-MAY-2016
Datum / Date: