



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2800AZ

TECHNOTHROMBIN<sup>®</sup> TGA Control Low 5x1mL

REF 5006330

LOT 2863B00.01

⌚ 31-Oct-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1mL	TGA Control Low	2863B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F61B00	63 - 147	128.3	OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D62B00	57 - 143	112.3	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

24. NOV. 2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

24. NOV. 2016  
Datum / Date: