



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2B00AZ.02

Ceveron TGA Control Low 5x1mL

REF 5006332  
LOT 2B66B00.01  
⌚ 30-Sep-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
pH	7.4 – 7.8	7.7	OK	
Verwendbarkeit: Usability: Ceveron® alpha	Peak Thrombin nM	78 - 130	121.7	OK
	Lag time min	3.6 – 6.0	4.4	OK
	AUC nM	863 - 1438	1121.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

25. OKT. 2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

25. OKT. 2016  
Datum / Date: