



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQ0D00AZ

Ceveron TGA CAL Set f. Ceveron

REF 5006347

LOT 0D64B00.01

31-Mar-2018

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA Buffer	9P31B02	31-Jul-2018
1x1mL	TGA Reaction Buffer	9K61B	31-Jul-2018
1x0.5mL	TGA Thrombin Standard	5T61B	31-Mar-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration Gegen / against WHO 01/580		800 – 900 nM	841	OK
Verwendbarkeit / usability Gerät / Analyzer: Ceveron: 10091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.990	1.0	OK
Kontrollwiederfindung Recovery of controls TGA RC High f. Ceveron 7F52B	TGA Control High (nM Peak Thrombin) LOT 1B52B00	66 – 310	203.6	OK
	TGA Control High (Lag time min) LOT 1B52B00	1.7 – 2.8	2.3	OK
	TGA Control High (AUC) LOT 1B52B00	1078 – 2845	2100.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

technoclone



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0D00AZ

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

23. SEP. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

23. SEP. 2016

Datum / Date: