



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0300.01

Technoview® Rivaroxaban CAL Set High 5x1mL

REF 5090171

LOT 0363B00.02

⌚ 31-Jan-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	Technoview Rivaroxaban CAL1 High	5161B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL2 High	5261B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL3 High	5361B00	31-Jan-2018
1x1mL	Technoview Rivaroxaban CAL4 High	5411B03	30-Jun-2018
1x1mL	Technoview Rivaroxaban CAL5 High	5561B00	31-Jan-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control High LOT 3351B00	242.6 – 404.3 ng/mL	370.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lucas ROSENMAYR

19. DEZ. 2016

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

20. DEZ. 2016

Datum / Date: