
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3400.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control low 5x1mL

REF 5090172
LOT 3451B00.02
 28-Feb-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		53.9 – 80.9 ng/mL	67.4	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	53.9 – 80.9 ng/mL	57.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von
Prepared by


Monika WEINBAUER

12. OKT. 2015

Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI


Datum / Date: