

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0F00AZ.03
---	---	-------------

**F VIII Inhibitor Reagent Kit 2-4T.**

**REF**            5152005  
**LOT**            0F62B00.02  
            31-Oct-2017

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
2x3.27mL	FVIII normal Plasma	1G61B00	31-Mar-2018
1x1mL	FVIII Inhibitor Plasma	2K52B00	31-Oct-2017
1x1mL	FVIII Inhibitor Free Plasma	2M61B0D	31-Mar-2018
1x17mL	Imidazol Puffer	9J49BA0	31-Oct-2019

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
<b>Verwendbarkeit Usability</b> FVIII clotting, Daptin TC Ceveron® alpha	FVIII Inhibitor Plasma <b>LOT</b> 2K52B00 13.5 – 22.5 BE/mL	15.4	OK
	FVIII Inhibitor Free Plasma <b>LOT</b> 2M61B0D < 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von  
Prepared by**

  
**Lucas ROSENMAYR**  


---

21. JULI 2016  
Datum / Date:

**Freigabe von  
Approved by**

  
**Michael KAFKA**  


---

21. JULI 2016  
Datum / Date: