



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4800AZ.01

F VIII Mangelplasma nativ, 50x1mL

REF 5154016
LOT 4865B00.01
⌚ 30-Sep-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P37B01	65.3 – 108.8 %	77.5	OK
	CCA LOT 3P38B01	15.0 – 25.0 %	19.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Suzana GRESAK PODAKAR

24. NOV. 2016

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

24. NOV. 2016

Datum / Date: