
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2K00.02
---	---	-----------

FVIII Inhibitorplasma HCV neg. 5x1mL

REF 5159010
LOT 2K51B0D.07
 31-May-2017


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.4	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII Inhibitor clotting, Daptin TC	14.1 – 23.5 BE/mL	16.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

07. NOV. 2016
Datum / Date:

Freigabe von
Approved by


Christoph ENNE

07. NOV. 2016
Datum / Date: