



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0Z00AZ.02

TECHNOCHROM[®] FXIII 3 x 3mL

REF 5360010
LOT 0Z66B00.02
⌚ 31-Dec-2017

Inhalt	Reagenz	LOT	⌚
3 x 3 mL	FXIII Activator Reagent	1873100/1	31-Dec-2017
3 x 3 mL	FXIII Detection Reagent	1873100/2	31-Dec-2017
3 x 3 mL	FXIII NADPH	1873100/3	31-Dec-2017
3 x 1 mL	FXIII Inhibitor Reagent	1873100/4	31-Dec-2017
1 x 6 mL	Stabilizer solution	1873100/5	31-Dec-2017

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B01	60.8% – 101.3%
	Coagulation Control A Lot: 3P38B01	28.5% – 47.5%

Durchführung von
Prepared by

Michael KAFKA 

19. JAN. 2017
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

19. JAN. 2017
Datum / Date: