
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9J00AZ.02
---	---	-------------

### Imidazol Puffer 25 mL

REF      5410008  
LOT      9J54BC0.01  
      30-Sep-2020

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.2 – 7.6	7.5	OK
Verwendbarkeit, Fibrinogen, Clauss method Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	2.52 – 3.78 g/L	3.15 g/L    OK
Usability, Fibrinogen, Clauss method Ceveron® alpha	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	0.93 – 1.39 g/L	1.11 g/L    OK
Etiketten Überprüfung Label Verification		OK	

**Durchführung von**  
**Prepared by**

Lucas ROSENMAYR

07-Oct-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI

07-Oct-2015  
Datum / Date: