



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7C00AZ

TECHNOTHROMBIN<sup>®</sup> TGA RB 5x0.5mL

REF 5006209

LOT 7C51B00.02

⌚ 31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
5x0.5 mL	TGA RB Reagenz	7C51B00	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00)	140 - 417	259.8	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

23. SEP. 2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

23. SEP. 2016  
Datum / Date: